

LETTER

Accurate predictions of iron redox state in silicate glasses: A multivariate approach using X-ray absorption spectroscopy

M. DARBY DYAR^{1,*}, MOLLY McCANTA², ELLY BREVES¹, C.J. CAREY³, AND ANTONIO LANZIROTTI⁴

¹Department of Astronomy, Mount Holyoke College, South Hadley, Massachusetts 01075, U.S.A.

²Department of Earth and Ocean Sciences, Tufts University, Medford, Massachusetts 02159, U.S.A.

³School of Information and Computer Sciences, University of Massachusetts at Amherst, Amherst, Massachusetts 01003, U.S.A.

⁴Center for Advanced Radiation Sources, University of Chicago, 5640 S. Ellis Avenue, Chicago, Illinois 60637, U.S.A.

ABSTRACT

Pre-edge features in the *K* absorption edge of X-ray absorption spectra are commonly used to predict Fe³⁺ valence state in silicate glasses. However, this study shows that using the entire spectral region from the pre-edge into the extended X-ray absorption fine-structure region provides more accurate results when combined with multivariate analysis techniques. The least absolute shrinkage and selection operator (lasso) regression technique yields %Fe³⁺ values that are accurate to ±3.6% absolute when the full spectral region is employed. This method can be used across a broad range of glass compositions, is easily automated, and is demonstrated to yield accurate results from different synchrotrons. It will enable future studies involving X-ray mapping of redox gradients on standard thin sections at 1 × 1 μm pixel sizes.

Keywords: XANES, EXAFS, silicate glass, lasso, PLS, multivariate analysis, redox state, ferric iron